

QUALITY CONTROL RANGES
MILLIPLEX[®] MAP Rat Myokine Magnetic Bead Panel
Catalog # RMYMAG-88K
Control Catalog # RMY-6088 Lot# RMY-102A.202A



Hormone	QC Level	Expected Range	Units	Hormone	QC Level	Expected Range	Units
BDNF	Control 1	118 - 245	pg/mL	FSTL-1	Control 1	1696 - 3523	pg/mL
	Control 2	918 - 1906	pg/mL		Control 2	14810 - 30759	pg/mL
EPO	Control 1	403 - 837	pg/mL	MSTN/GDF8	Control 1	23 - 48	ng/mL
	Control 2	3667 - 7616	pg/mL		Control 2	191 - 396	ng/mL
IL-15	Control 1	10 - 21	ng/mL	IRISN	Control 1	7.4 - 15	ng/mL
	Control 2	84 - 175	ng/mL		Control 2	66 - 137	ng/mL
FGF21	Control 1	74 - 153	pg/mL	LIF	Control 1	734 - 1525	pg/mL
	Control 2	658 - 1366	pg/mL		Control 2	6792 - 14107	pg/mL
Fractalkine	Control 1	90 - 187	pg/mL	Osterocrin	Control 1	165 - 344	pg/mL
	Control 2	814 - 1691	pg/mL		Control 2	1350 - 2804	pg/mL
IL-6	Control 1	4.2 - 8.7	ng/mL	SPARC	Control 1	3.4 - 7.1	ng/mL
	Control 2	35 - 73	ng/mL		Control 2	33 - 69	ng/mL

Note: The Quality Control Ranges are generated with overnight assay format using serum matrix provided in the kit. Quality Control values in culture media are not tested.

RMY-6088: RMY-102A.202A (12 May 2021)

• EMD MILLIPORE • 1-800-645-5476 • 1-800-645-5439 (FAX) • www.emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA.