

Assurance® GDS MPX Top 7 STEC

NF Validation Certificate N° TRA 02/13-04/22

Part No: 71015-100 (100 tests)

General Description

Assurance® GDS MPX for Top 7 STEC is an automated nucleic acid amplification system for the detection of *E. coli* O157:H7 and the "Top Six" non-O157 Shiga Toxigenic *E. coli* (STEC) in raw beef meat and ready-to-cook beef meat products, raw milk and dairy products. The Top Six non-O157 STEC are defined as *E. coli* belonging to serogroups O103, O111, O121, O145, O26 or O45 that possess both the *eae* gene and at least one of the Shiga toxin genes, *stx1* or *stx2*. Assurance® GDS MPX for Top 7 STEC utilizes a proprietary IMS-based sample preparation procedure to capture organisms belonging to 7 specific Top STEC O-serogroups (O103, O111, O121, O145, O26, O45, and O157) prior to genetic analysis for the associated pathogenicity genes. ISO/TS 13136 (2012) targets only the Top 5 STEC (O157, O103, O111, O145, and O26). This NF method is validated only for the detection and confirmation of the Top 5 STEC serotypes.

Kit Components

Each Assurance® GDS MPX Top 7 STEC kit contains the following:

- MPX Top STEC Amplification Tubes
- Top 7 STEC Concentration Reagent
- Resuspension Buffer Tq
- Top STEC Wash Solution

Equipment / Materials Required

Other necessary materials not provided include:

- mEHEC® media
- Brain heart infusion (BHI) broth
- Assurance® GDS Rotor-Gene® thermocycler
- GDS rotor and locking ring
- Laptop computer and software v2.3.103
- PickPen® device and PickPen® tips
- Vortex mixer (IKA® MS3 or equivalent)
- Adhesive film strips
- Sample wells and sample well base
- Resuspension plate
- Stomacher® paddle homogenizer or equivalent
- Stomacher®-type bags with filter or equivalent
- 8-channel micropipette capable of accurately dispensing 30 µL
- Adjustable micropipette capable of accurately dispensing 1.0 mL
- Repeat pipette
- Repeat pipette tips (0.5 mL and 10 mL)

Filter barrier micropipette tips (50 µL and 1.0 mL)

Gel cooling block

Incubator capable of maintaining 41.5 ± 1 °C

Incubator capable of maintaining 37 ± 1 °C

Freezer capable of maintaining -20 ± 5 °C

Refrigerator capable of maintaining 5 ± 3 °C

Enrichment Media Preparation

- A. For 25 g sample, pre-warm 225 mL sterile deionized water at 41.5 °C overnight. On day of use, aseptically transfer 7.1 g of mEHEC® media into the pre-warmed sterile water. Gently mix to rehydrate the media. Use prepared broth within 6 h.
- B. For 375 g sample, pre-warm 1500 mL sterile deionized water at 41.5 °C overnight. On day of use, aseptically transfer 47.3 g of mEHEC® media into the pre-warmed sterile water. Gently mix to rehydrate the media. Use prepared broth within 6 h.
- C. Alternatively, mEHEC® media can be prepared in advance and autoclaved. Add 31.6 g media per liter of deionized water. Stir to rehydrate the media, dispense into desired volume and autoclave at 121 °C for 15 min. Broth must be pre-warmed to 41.5 °C overnight prior to sample addition.

Sample Preparation

Application in raw, frozen, ready-to-cook (25 – 375 g) and seasoned (25 g only) beef meat products, raw milk and dairy products.

Note: For this method, when a temperature of 41.5 °C is specified, the acceptable temperature range is 41.5 ± 1 °C.

A. Test Portion Preparation & Enrichment

Please see APPENDIX A for Enrichment Methods Table

1. **Raw beef meat** – For 25 g test portion, add sample into 225 mL pre-warmed (41.5 °C) mEHEC® media. For 375 g test portion, weigh into 1,500 mL pre-warmed (41.5 °C) mEHEC® media. Homogenize or mix sample for 2 min. Incubate 25 g samples for 8 – 16 h at 41.5 °C. Incubate 375 g samples for 10 – 18 h at 41.5 °C. For frozen state beef, extend incubation time by 2 h.
2. **Raw milk and dairy (including raw cheese) products** – For 25 g (25 mL) test portion, weigh into 225 mL of pre-warmed (41.5 °C) mEHEC® media. Homogenize or mix sample for 2 min. Incubate 25 g samples for 18 – 24 h at 41.5 °C.

Note: Contact Technical Services (BioMTS@milliporesigma.com) for recommended procedures for testing alternate sample types or sizes.

B. Sample Extraction Protocol

Change gloves prior to handling reagents.

1. All Foods except Cheese from Raw Milk

- a. Vortex **Top 7 STEC Concentration Reagent**. Immediately transfer 20 µL to each of the required number of GDS sample wells (1 well/sample) using a repeat pipette and a 0.5 mL pipette tip. Cover sample wells with adhesive film strips.
- b. Transfer 1.0 mL of **Top STEC Wash Solution** to each of 2 additional GDS sample wells (2 wells/sample) using a repeat pipette and a 10 mL pipette tip. Cover sample wells with adhesive film strips.

- c. Transfer 45 μ L of **Resuspension Buffer Tq** to the sample wells in the resuspension plate using a repeat pipette and a 0.5 mL pipette tip. Cover resuspension plate with adhesive film strips.
- d. Carefully remove the adhesive film from 1 strip of sample wells containing Top 7 STEC Concentration Reagent. Mix incubated sample enrichments. Add 1.0 mL of enrichment to each sample well. Avoid transferring food particles. A new pipette tip must be used for each sample. Cover each strip of sample wells with a new adhesive film prior to adding enrichments to the next strip of wells. **Immediately return samples to incubator for use during confirmation, if necessary.**

Note: Enrichments for raw milk and dairy products, after 18 h, may be retained at room temperature during sample analysis.

- e. Place sealed sample wells containing the Top 7 STEC Concentration Reagent and enrichments on the vortex mixer and vortex at approximately 900 rpm for 10–20 min. If necessary, adjust rpm to be certain that liquid does not contact adhesive film.
- f. Carefully remove and discard the adhesive film from 1 strip of enrichments. Remove the corresponding adhesive film from 2 strips of sample wells containing Top STEC Wash Solution.
- g. Load tips onto the PickPen[®] device, ensuring that the tips are firmly in place on the PickPen[®] tool. Extend the PickPen[®] magnets and insert tips into the strip of enrichments.

Stir tips gently for 30 s while continually moving up and down from the surface to the bottom of the well. Gently tap the PickPen[®] tips against the side of the sample wells to remove excess media droplets.
- h. Transfer PickPen[®] tips to corresponding first set of sample wells containing Top STEC Wash Solution and retract PickPen[®] magnets to release particles into Top STEC Wash Solution.
- i. Discard PickPen[®] tips and load a new set of tips onto the PickPen[®] device.
- j. Extend the PickPen[®] magnets and insert tips into the strip of wells containing the Top STEC Wash Solution and particles. Stir tips gently for 30 s while continually moving up and down from the surface to the bottom of the well. Gently tap the PickPen[®] tips against the side of the sample wells to remove excess droplets of Wash Solution.
- k. Transfer PickPen[®] tips to the second set of sample wells containing fresh Top STEC Wash Solution and gently swirl for 10 s (do not release particles into solution). Tap the PickPen[®] tips against the side of the sample wells to remove excess droplets of Top STEC Wash Solution.
- l. Remove the adhesive film from resuspension plate. Transfer particles to corresponding row of the prepared resuspension plate. With tips submerged, retract the PickPen[®] magnets and tap gently to release particles into the Resuspension Buffer Tq. Cover resuspension plate with adhesive film.
- m. Repeat steps (f) through (l) for all samples using new tips for each strip of enrichments.

PROCEED TO TEST PROCEDURE SECTION

2. Cheese from Raw Milk

- a. Vortex **Top 7 STEC Concentration Reagent**. Immediately transfer 20 μ L to each of the required number of GDS sample wells (1 well/sample) using a repeat pipette and a 0.5 mL pipette tip. Cover sample wells with adhesive film strips.
- b. Transfer 1.0 mL of **Top STEC Wash Solution** to each of 2 additional sample wells (2 wells/sample) using a repeat pipette and a 10 mL pipette tip. Cover sample wells with adhesive film strips.
- c. Dispense 0.5 mL of sterile Brain Heart Infusion (BHI) broth to sample wells (1 well / sample) using a repeat pipette and a 10 mL pipette tip. Cover sample wells with adhesive film strips.
- d. Transfer 45 μ L of **Resuspension Buffer Tq** to the sample wells in the resuspension plate using a repeat pipette and a 0.5 mL pipette tip. Cover resuspension plate with adhesive film strips.
- e. Carefully remove the adhesive film from 1 strip of sample wells containing Top 7 STEC Concentration Reagent. Mix incubated enrichments. Add 1.0 mL of incubated enrichment to each sample well. Avoid transferring food particles. A new pipette tip must be used for each sample. Cover each strip of sample wells with a new adhesive film prior to adding samples to a new strip of wells. **Immediately return samples to incubator for use during confirmation, if necessary.**

Note: Enrichments for raw cheese, after 18 h, may be retained at room temperature during sample analysis.

- f. Place sealed sample wells containing the Top 7 STEC Concentration Reagent and enrichments on the vortex mixer and agitate at approximately 900 rpm for 10–20 min. If necessary, adjust rpm to be certain that liquid does not contact adhesive film.
- g. Carefully remove and discard the adhesive film from 1 strip of enrichments. Remove the corresponding adhesive film from 1 strip of sample well containing Top STEC Wash Solution. Also remove the corresponding adhesive film for strip of sample wells containing BHI.
- h. Load tips onto the PickPen® device, ensuring that the tips are firmly in place on the PickPen® tool. Extend the PickPen® magnets and insert tips into the first strip of enrichments.
Stir tips gently for 30 s while continually moving up and down from the surface to the bottom of the well. Gently tap the PickPen® tips against the side of the sample wells to remove excess media droplets.
- i. Transfer PickPen® tips to set of sample wells containing fresh Top STEC Wash Solution and gently swirl for 10 s (do not release particles into solution). Tap the PickPen® tips against the side of the sample wells to remove excess droplets of Top STEC Wash Solution.
- j. Transfer PickPen® tips to corresponding sample wells containing BHI and retract PickPen® magnets to release particles into BHI.
- k. Cover each strip of samples wells containing BHI with an adhesive film prior to adding additional samples to a new strip of sample wells. Incubate BHI sample wells containing particles for 2 – 4 h at 37 ± 1 °C. Repeat steps (g) – (k) for all samples.
- l. After incubation, remove the corresponding adhesive film from remaining strip of sample well containing Top STEC Wash Solution. Carefully remove and discard the adhesive film from strip of BHI samples. Extend the PickPen® magnets and insert tips into the strip of sample wells containing the BHI and particles.
Stir tips gently for 30 s while continually moving up and down from the surface to the bottom of the well. Gently tap the PickPen® tips against the side of the sample wells to remove excess droplets of BHI.
- m. Transfer PickPen® tips to the second set of sample wells containing fresh Top STEC Wash Solution and gently swirl for 10 s (do not release particles into solution). Tap the PickPen® tips against the side of the sample wells to remove excess droplets of Top STEC Wash Solution.
- n. Transfer particles to corresponding row of the prepared resuspension plate. With tips submerged, retract the PickPen® magnets and tap gently to release particles into the Resuspension Buffer Tq. Cover resuspension plate with adhesive film.
- o. Repeat steps (l) through (n) for all samples using new PickPen® tips for each strip of enrichments.

Test Procedure (Amplification & Detection)

Change gloves prior to handling reagents.

A. Preparation of Gel Cooling Block

1. Prior to initial use, the gel cooling block must be stored in the freezer (-20 ± 5 °C) for minimum 6 h. When frozen the gel cooling block will change color from pink to purple. When not in use, the gel cooling block should continue to be stored at -20 ± 5 °C.
2. Between each use the gel cooling block should be returned to the freezer until it has turned completely purple, indicating it is ready for use. This may take up to 2 h.

B. Preparation of Amplification Tubes

1. The Assurance® GDS Rotor-Gene® set-up and data entry should be completed prior to transferring samples from the resuspension plate into the **Amplification Tubes**.
2. Remove MPX Top STEC Amplification Tubes from foil pouch and place them in the frozen gel cooling block. Reseal pouch.
3. Open Amplification Tubes. Using a multi-channel pipette and filter barrier tips, briefly pipette up and down the Resuspension Buffer Tq to mix beads in resuspension plate wells. Transfer 30 µL of sample from the resuspension plate wells into each Amplification Tube. Firmly press down on each Amplification Tube lid to close. Visually inspect each tube to ensure that the cap is securely sealed.
4. Place Amplification Tubes into Assurance® GDS Rotor-Gene® in sequential order, beginning with position #1. Start Rotor-Gene® cycle. Refer to Assurance® GDS User Manual for detailed instructions on operating the Rotor-Gene® thermocycler.

Note: The Assurance® GDS Rotor-Gene® must be started within 20 min after addition of the samples to the Amplification Tubes.

Results

Upon completion of the run, the Assurance® GDS Rotor-Gene® software will provide a results table. Each sample will be identified as **Positive** or **Negative** for Top STEC, **Positive** or **Negative** for *E. coli* O157:H7, or **No Amp**. The individual gene results (*eae*, *stx1*, *stx2*) are also presented.

Top STEC (*eae/stx*) Results:

Positive: Samples are presumptive positive for Top STEC, meaning they are *E. coli* that belong to O-serogroups O103, O111, O121, O145, O26, O45, and O157 and contain the *eae* gene and one or both of the Shiga toxin genes, *stx1* or *stx2*.

Negative: Samples are negative for Top STEC.

No Amp: Amplification did not occur. Repeat the test beginning from step **B. Sample Extraction Protocol**. If the No Amp result is repeated, contact Technical Services (BioMTS@milliporesigma.com).

No.	Name	Top STEC Result	eae Result	stx1 result	stx2 result	Assay	Kit lot
1	Sample 1	Positive	+	+	+	Top STEC MPX	abc123
2	Sample 2	Positive	+	+	-	Top STEC MPX	abc123
3	Sample 3	Positive	+	-	+	Top STEC MPX	abc123
4	Sample 4	Negative	+	-	-	Top STEC MPX	abc123
5	Sample 5	Negative	-	+	+	Top STEC MPX	abc123
6	Sample 6	Negative	-	+	-	Top STEC MPX	abc123
7	Sample 7	Negative	-	-	+	Top STEC MPX	abc123
8	Sample 8	Negative	-	-	-	Top STEC MPX	abc123
9	Sample 9	No Amp	-	-	-	Top STEC MPX	abc123

E. coli O157:H7 Results:

Positive: Samples are presumptive positive for *E. coli* O157:H7. Certain strains of STEC belonging to the O145 serogroup may also be indicated as positive for *E. coli* O157:H7.

Negative: Samples are negative for *E. coli* O157:H7.

No Amp: Amplification did not occur. Repeat the test beginning from step **B. Sample Extraction Protocol**. If the No Amp result is repeated contact Technical Services (BioMTS@milliporesigma.com).

No.	Name	<i>E. coli</i> O157:H7 Result	Assay	Kit lot
1	Sample 1	Positive	Top STEC MPX	abc123
2	Sample 2	Positive	Top STEC MPX	abc123
3	Sample 3	Positive	Top STEC MPX	abc123
4	Sample 4	Negative	Top STEC MPX	abc123
5	Sample 5	Negative	Top STEC MPX	abc123
6	Sample 6	Negative	Top STEC MPX	abc123
7	Sample 7	Negative	Top STEC MPX	abc123
8	Sample 8	Negative	Top STEC MPX	abc123
9	Sample 9	No Amp	Top STEC MPX	abc123

Note: Assurance® GDS MPX for Top 7 STEC is intended for detection of the Top 7 STEC serotypes; however, any *eae*-positive *E. coli* isolate that also contains *stx1* and/or *stx2* would be considered a potentially pathogenic strain.

Confirmation

Note: For raw milk cheese samples, confirmation must proceed from the mEHEC® enrichment. Confirmation cannot proceed from the BHI subculture.

- A. An aliquot of the mEHEC® enrichment from GDS MPX for Top 7 STEC positive samples may be culturally confirmed for Top 7 STEC via ISO/TS 13136:2012 *Microbiology of food and animal feed — Real-time*

polymerase chain reaction (PCR)-based method for the detection of food-borne pathogens — Horizontal method for the detection of Shiga toxin-producing Escherichia coli (STEC) and the determination of O157, O111, O26, O103 and O145 serogroups.

- B. Top 6 non-O157 STEC, Top 6 non-O157 STEC are identified by a combination of a positive result from TOP STEC channel and a negative result from O157 channel. An aliquot of the mEHEC[®] enrichment from GDS MPX for Top 7 STEC positive samples may be confirmed for Top 6 STEC via secondary analysis by the MPX ID assay.

Note: The NF method will be validated only for the confirmation of the Top 4 non-O157 STEC (O111, O26, O103 and O145) serotypes listed in ISO/TS 13136 (2012).

- C. E. coli O157:H7, an aliquot of the mEHEC[®] enrichment from GDS MPX for Top 7 STEC positive samples may be confirmed for *E. coli* O157:H7 via secondary analysis by GDS EHEC ID and MPX ID assays, to distinguish between O157 and O145, as needed. Certain strains of STEC belonging to the O145 serogroup may also be indicated as positive for *E. coli* O157:H7.

Storage

Store Assurance[®] GDS MPX for Top 7 STEC kit components at 5 ± 3 °C. Kit expiration is provided on the product box label.

Precautions

Assurance[®] GDS MPX for Top 7 STEC kit must be used as described herein.

Do not use Assurance[®] GDS MPX for Top 7 STEC reagents that have expired. Do not use test kit beyond expiration date on the product box label.

Safety

Assurance[®] GDS MPX for Top 7 STEC kit.—This product is not intended for human or veterinary use. Assurance[®] GDS MPX for Top 7 STEC must be used as described in the package insert. Contents of the test may be harmful if swallowed or taken internally. The user should read, understand and follow all safety information in the instructions for the Assurance[®] GDS for MPX Top 7 STEC kit. Retain the safety instructions for future reference.

Do not open or autoclave used Amplification Tubes. —After run is complete, place used Amplification Tubes into a sealed container with sufficient volume of a 10% bleach solution to cover tubes for a minimum of 15 min or double bag amplification tubes and dispose outside of the lab. Follow all applicable local, state/provincial, and/or national regulations on disposal of amplification tubes. If contamination is suspected, moisten paper towel with bleach solution and wipe all lab benches and equipment surfaces with 10% bleach solution. Avoid spraying bleach solution directly onto surfaces. Allow bleach solution to remain on surfaces for a minimum of 15 min before wiping clean with 70% isopropyl alcohol solution.

To prepare 10% bleach solution, add 10 mL of commercially available bleach containing at least 5% sodium hypochlorite to 90 mL of deionized water. The minimum final concentration of sodium hypochlorite in the bleach solution should be 0.5%. The bleach solution is stable for 7 days from preparation. To prepare 70% isopropyl alcohol solution, add 70 mL of pure isopropyl alcohol to 30 mL of deionized water or buy commercially available 70% isopropyl alcohol.

Assurance[®] GDS Rotor-Gene[®].—Improper use of the Assurance[®] GDS Rotor-Gene[®] may cause personal injuries or damage to the instrument. Some components may pose a risk of personal injury due to excessive heat if improperly handled. For safe use, the instrument must only be operated by qualified laboratory personnel who have been appropriately trained. Servicing of instrument must only be performed by our Service Engineers.

Sample Enrichment.—To reduce the risks associated with exposure to chemicals and biohazards, perform pathogen testing in a properly equipped laboratory under the control of trained personnel. Always follow standard laboratory safety practices, including wearing appropriate personal protective apparel and eye protection, PPE, while handling reagents and contaminated samples. Avoid contact with the contents of the enrichment media and reagent tubes after amplification. Dispose of enriched samples according to current industry standards. Decontaminate and dispose of materials in accordance with good laboratory practices and in accordance with local, state, and federal regulations.

Shiga Toxigenic *E. coli* (STEC) Precautions.—STEC are a biosafety level-3 organism. Biological samples, such as enrichments, have the potential to transmit infectious diseases. Follow all applicable local, state/provincial, and/or national regulations on disposal of biological wastes. Wear appropriate protective equipment which includes, but is not limited to, protective eyewear, face shield, clothing/laboratory coat, and gloves. All work should be conducted in properly equipped facilities utilizing the appropriate safety equipment (for example, physical containment devices). Individuals should be trained in accordance with applicable regulatory and company/institution requirements before working with potentially infectious materials. All enrichment broths should be sterilized following any culture based confirmatory steps. Clean the workstations and laboratory equipment with a disinfectant of choice before and after lab activities (sodium hypochlorite solution, phenol solution, quaternary ammonium solution, etc.).

APPENDIX A - Enrichment Methods

Table 1. Sample Type and Enrichment Method for Top 7 STEC

Food Category	Media	Sample size	Sample:Media Ratio (Media Volume)	Enrichment Time	Incubation Temperature
No BHI subculture					
Raw beef meats and ready-to-cook beef meat products	mEHEC®	25 g	1:10 (225 mL)	8-16 h	41.5 ± 1 °C
		375 g	1:5 (1500 mL)	10-18 h	
Seasoned beef meats	mEHEC®	25 g	1:10 (225 mL)	8-16 h	41.5 ± 1 °C
Raw milk and dairy products (except raw milk cheese)	mEHEC®	25 g	1:10 (225 mL)	18-24 h	41.5 ± 1 °C
With BHI subculture					
Raw milk cheese	mEHEC®	25 g	1:10 (225 mL)	18-24 h + 2 - 4 h BHI	41.5 ± 1 °C

NF Validation certificate granted by AFNOR Certification for TRANSIA® PLATE Listeria as an alternative method of analysis for all food products and industrial production environmental samples in relation to the reference method described in the ISO EN 11290-1 international standard in accordance with EN ISO 16140-2 (2016). For more information about the NF VALIDATION certification, please refer to the certificate available at www.afnor-validation.com



TRA 02/13-04/22

ALTERNATIVE METHODS OF ANALYSIS FOR AGRIBUSINESS

www.afnor-validation.org

Manufacturing Entity

BioControl Systems, Inc, 12822 SE 32nd St, Bellevue, WA 98005, USA.

Tel: 425-586-3300 www.sigmaldrich.com

BioControl Systems, Inc is an affiliate of Merck KGaA, Darmstadt, Germany.

MilliporeSigma
400 Summit Drive
Burlington, MA 01803
USA

© 2022 Merck KGaA, Darmstadt Germany and/or its affiliates. All Rights Reserved. MilliporeSigma, Millipore, the Vibrant M Assurance GDS, mEHEC, PickPen, Transia, and Rotor-Gene are trademarks of Merck KGaA, Darmstadt Germany or its affiliate. All other trademarks are the property of their respective owners. Detailed information on trademarks is available via publicly accessible resources.

Lit Code: MS_DU9903EN Ver. 1.0
10/2022
55410-R000

