



Product Information


SIGMA QUALITY CONTROL TEST PROCEDURE

Enzymatic Assay of ESTERASE

(EC 3.1.1.1)

Sigma Prod. No. E-2884, E-3019, and E-3128

PRINCIPLE:



CONDITIONS: T = 25°C, pH = 8.0

METHOD: Titrimetric

REAGENTS:

- A. 10 mM Borate Buffer, pH 8.0 at 25°C
(Prepare 500 ml in deionized water using Boric Acid, Sigma Prod. No. B-0252. Adjust to pH 8.0 with 1 M NaOH.)
- B. 0.1% (v/v) Ethyl Butyrate Solution
(Prepare 200 ml in Reagent A using Butyric Acid Ethyl Ester, Sigma Prod. No. B-2391.)
- C. Ethyl Butyrate
(Use Butyric Acid Ethyl Ester, Sigma Prod. No. B-2391.)
- D. 10 mM Sodium Hydroxide Solution-Standardized (NaOH)
(Prepare 50 ml in cold deionized water using Sodium Hydroxide, Pellets, Sigma Prod. No. S-5881. Standardize according to the ACS Reagent Procedure.¹)
- E. Esterase Enzyme Solution
(Immediately before use, prepare a solution containing 50 units/ml of Esterase in cold Reagent A.)

PROCEDURE:

Using a suitable pH meter in conjunction with a magnetic stirrer, pipette (in milliliters) the following reagents into a suitable titration vessel:

	<u>Test</u>
Reagent B (Ethyl Butyrate Soln)	25.0

**Enzymatic Assay of ESTERASE
(EC 3.1.1.1)
Sigma Prod. No. E-2884, E-3019, and E-3128**

PROCEDURE: (continued)

Equilibrate to 25°C. Then add:

	<u>Test</u>
Reagent C (Ethyl Butyrate)	0.025

Adjust the pH to 8.1 using Reagent D (NaOH). Then add:

Reagent E (Enzyme Solution)	0.10
-----------------------------	------

Start a timer when the pH reaches 8.0. Run the reaction for 1-5 minutes. Maintain the pH of the reaction mix at pH 8.0 by the addition of small volumes (0.050 ml) of Reagent D. Record the volume of Reagent D used to maintain the pH at 8.0 and the time required to consume the added Reagent D.

CALCULATION:

$$\text{Units/ml enzyme} = \frac{(\text{Molarity of NaOH})(\text{NaOH})(1000)(\text{df})}{(\text{T})(0.10)}$$

NaOH = Volume (in milliliters) of Reagent D used in the assay

1000 = Conversion factor from millimoles to micromoles (Unit definition)

df = Dilution factor

T = Time required (in minutes) to consume the added Reagent D (NaOH) while maintaining the pH at 8.0

0.1 = Volume (in milliliter) of enzyme used

$$\text{Units/mg protein} = \frac{\text{units/ml enzyme}}{\text{mg protein/ml enzyme}}$$

UNIT DEFINITION:

One unit will hydrolyze 1.0 μmole of ethyl butyrate to butyric acid and ethanol per minute at pH 8.0 at 25°C.

INITIAL ASSAY CONCENTRATIONS:

In a 25.125 ml reaction mix, the initial concentrations are 10 mM borate, 0.2% (v/v) ethyl butyrate and 5 units esterase.

**Enzymatic Assay of ESTERASE
(EC 3.1.1.1)
Sigma Prod. No. E-2884, E-3019, and E-3128**

REFERENCES:

(1993) *Reagent Chemicals ACS Specifications*, 8th ed., p 95, American Chemical Society, Washington, DC

Adler, A.J. and Kistiakowsky, G.B. (1962) *Journal of the American Chemicals Society* **84**, 695-703

NOTES:

1. Standardization of NaOH is described in the (1993) *Reagent Chemical ACS Specification* reference.
2. This assay is based on the assay procedure described in Adler, A.J. and Kistiakowsky, G.B. (1962).
3. Where Sigma Product or Stock numbers are specified, equivalent reagents may be substituted.

Sigma warrants that the above procedure information is currently utilized at Sigma and that Sigma products conform to the information in Sigma publications. Purchaser must determine the suitability of the information and products for its particular use. Upon purchase of Sigma products, see reverse side of invoice or packing slip for additional terms and conditions of sale.