

QUALITY CONTROL RANGES



Milliplex Rat Pituitary Magnetic Kit Kit Catalog# RPTMAG-86K

Control Catalog # LRPT-6086 Lot# RPT-113 and RPT-213

<u>Hormone</u>	<u>QC Level</u>	<u>Expected Range</u>	<u>Units</u>
ACTH	Control I	32 - 66	pg/mL
	Control II	179 - 371	pg/mL
BDNF	Control I	38 - 78	pg/mL
	Control II	204 - 423	pg/mL
FSH	Control I	1156 - 2400	pg/mL
	Control II	6298 - 13081	pg/mL
GH	Control I	197 - 410	pg/mL
	Control II	1065 - 2213	pg/mL
LH	Control I	234 - 485	pg/mL
	Control II	1235 - 2564	pg/mL
Prolactin	Control I	602 - 1250	pg/mL
	Control II	3247 - 6743	pg/mL
TSH	Control I	154 - 320	pg/mL
	Control II	820 - 1703	pg/mL

Note: The Quality Control Ranges are generated with overnight assay format using serum matrix provided in the kit. Quality Control values in culture media are not tested.

LRPT-6086: RPT-113.213 (06 May 2021)

EMD Millipore • 1-800-645-5476 • 1-800-645-5439 (FAX) • emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA.