



Quality Control Ranges

MILLIPLEX® MAP Sepsis Magnetic Bead Panel 2 Kit Catalog # HSP2MAG-63K

Control Catalog # HSP2-6063-2, Lot# HSP2-110 and HSP2-210

<u>Analyte</u>	<u>QC Level</u>	<u>Expected Range</u>	<u>Units</u>
Granzyme B	Control I	19 - 40	pg/mL
	Control II	89 - 185	pg/mL
HSP70	Control I	12172 - 25281	pg/mL
	Control II	64500 - 133961	pg/mL
IL-1 α	Control I	93 - 194	pg/mL
	Control II	509 - 1057	pg/mL
IL-8	Control I	45 - 93	pg/mL
	Control II	242 - 502	pg/mL
MIP-1 α	Control I	73 - 151	pg/mL
	Control II	352 - 731	pg/mL
MIP-1 β	Control I	93 - 193	pg/mL
	Control II	471 - 978	pg/mL
MMP-8	Control I	1559 - 3238	pg/mL
	Control II	8794 - 18264	pg/mL

Note: The Quality Control Ranges are generated with overnight assay format using serum matrix provided in the kit. Quality Control values in culture media are not tested.